




# *DQEPro*

## Dynamic (Fluoroscopic) DQE Sample Report

- Clinical-environment test



This report summarizes the analysis of a new CsI-based flat-panel detector in a clinical environment. The dynamic DQE is evaluated operating the detector at 30 frames/sec pulsed fluoroscopy using an RQA-5 (70 kV) spectrum. The assessment was completed in 20 min and required the acquisition of five image sequences (5.5 to 13 sec each) corresponding to a total of approximately 1000 images that were exported in tiff format. The scatter-rejection grid was not in place.

1. Both x and y presampling MTF curves show excellent spatial resolution appropriate for 2x2 binning of detector elements. The MTF 50%, 10% and 5% frequencies were 0.9, 1.9 and 2.2 cycles/mm in both x and y directions.
2. The temporal NPS is nearly flat with temporal frequencies, and the temporal autocovariance function is a single sharp peak with very small tails. This indicates very little lag between frames. The dynamic DQE correction factor was 1.03, corresponding to a correction of only 3%.
3. The two-dimensional NPS shows smooth circular symmetry with no unexpected noise structure.
4. The zero-frequency DQE value is approximately 0.82.
5. System gain is 3.2 pixel value / incident x-ray quantum.
6. The exposure waveform report shows near-uniform exposure pulses at 30 /sec.

20150116-121856


**Study comment:** Sample study  
**Study date-time:** Fri 16 Jan 2015 12:18:56  
**Grid:**  
**X-ray spectrum:** Dynamic, RQA-5 (70 kV)  
**Set technique:** 70 kV, 200 mA  
**Incident image-plane exposure, air KERMA:** 0.061 mR, 0.54 uGy  
**Half-value layer:** 7.1 mmAl  
**Q<sub>o</sub>:** 2.64e+08 q/mm<sup>2</sup>/R, 3.02e+10 q/mm<sup>2</sup>/Gy  
**Pixel size in image plane, x y:** 388 x 388 um (measured by DQEPro)  
**Average dark and open image pixel values:** 1162.4 8408.4  
**Average system gain:** 0.4469 DV/(q/mm<sup>2</sup>); 118195 DV/mR; 2.96 DV/q  
**Dynamic DQE correction factor:** 0.97

**Approximate MTF and DQE Values:**




cy/mm	x-MTF	x-DQE	cy/mm	y-MTF	y-DQE
0.00	1.00		0.00	1.00	
0.25	0.90	0.77	0.25	0.90	0.77
0.50	0.76	0.71	0.50	0.77	0.71
0.75	0.60	0.63	0.75	0.61	0.63
1.00	0.45	0.50	1.00	0.46	0.50
1.25	0.32	0.33	1.25	0.33	0.33
1.50	0.22		1.50	0.23	

Warnings were generated that may affect accuracy of results. Please see Message Summary report for details.

**Image-x Direction**



**Image-y Direction**




20150116-121856

**Study comment:** Sample study  
**Study date-time:** Fri 16 Jan 2015 12:18:56  
**X-ray spectrum:** Dynamic, RQA-5 (70 kV)  
**Incident image-plane exposure:** 0.061 mR, 0.54 uGy  
**Window width:** 100 mm  
**Pixel size in image plane:** 388 um (measured by DQEPro)  
**Sampling cut-off frequency:** 1.3 cycles/mm  
**MTF 50%, 10% and 5% frequencies:** 0.9 1.9 2.2 cycles/mm


**Approximate MTF Values:**

cycles/mm	MTF
0.00	1.00
0.25	0.90
0.50	0.76
0.75	0.60
1.00	0.45
1.25	0.32
1.50	0.22
1.75	0.14
2.00	0.08
2.25	0.04
2.50	0.01
2.75	0.01
3.00	0.03


**Edge-Spread Function**



**Line-Spread Function**



**Modulation Transfer Function**




20150116-121856

**Study comment:** Sample study  
**Study date-time:** Fri 16 Jan 2015 12:18:56  
**X-ray spectrum:** Dynamic, RQA-5 (70 kV)  
**Incident image-plane exposure:** 0.061 mR, 0.54  $\mu$ Gy  
**Window width:** 100 mm  
**Pixel size in image plane:** 388  $\mu$ m (measured by DQEPro)  
**Sampling cut-off frequency:** 1.3 cycles/mm  
**MTF 50%, 10% and 5% frequencies:** 0.9 1.9 2.2 cycles/mm


**Approximate MTF Values:**

cycles/mm	MTF
0.00	1.00
0.25	0.90
0.50	0.77
0.75	0.61
1.00	0.46
1.25	0.33
1.50	0.23
1.75	0.15
2.00	0.09
2.25	0.04
2.50	0.01
2.75	0.02
3.00	0.03


**Edge-Spread Function**



**Line-Spread Function**



**Modulation Transfer Function**




20150116-121856

**Study comment:** Sample study  
**Study date-time:** Fri 16 Jan 2015 12:18:56  
**X-ray spectrum:** Dynamic, RQA-5 (70 kV)  
**Incident image-plane exposure:** 0.061 mR, 0.54 uGy  
**Pixel size in image plane:** 388 x 388 um (measured by DQEPro)  
**Average dark and open pixel values:** 1162.4 8408.4  
**Average dark-subtracted open pixel value:** 7245.9  
**Lag fraction (approximate):** 0.07  
 Lag fraction estimate assumes single-exponential lag only.  
**Dynamic NPS zero value:** 1.01  
**Dynamic NPS quantum-correction factor:** 1.03


Normalized NPS has been corrected for lag and scaled by  $(\text{incident-q} / \text{mm}^2) / (\text{mean-pixel-value})^2$  such that an ideal detector has unity NPS.

In the NPS plots below, blue markers indicate total image noise while green markers indicate detector readout noise. Images are quantum-noise limited only when readout noise is a small fraction of total image noise.


**Two-Dimensional NPS**




**Temporal Autocovariance**




**Temporal NPS**



**Image-x Direction**








**Image-y Direction**



20150116-121856






Data folder: /Users/DQE/data

These waveforms show relative x-ray intensity profiles measured by DQEPro for each exposure. Green line segments correspond to identified exposure pulses. Leading and trailing baselines are shown as dark blue segments. The first pulse should start at 0 sec. Ensure pulses are correctly identified and free of undesirable artifacts that may indicate problems with the x-ray generator or tube. Double-click on any waveform to open a review window.



20150116-121856


These waveforms show average pixel-value profiles in the acquired images for each exposure. Green line segments correspond to images selected for the dynamic analysis. The number of images used with Calib and Open type exposures will be the largest possible power of two. Double-click on any plot to open a review window.




20150116-121856

**Detector:**  
**Detector SN:**  
**Study comment:** Sample study  
**Study date-time:** Fri 16 Jan 2015 12:18:56  
**Potential defects include real defects and pixels obscured from x-ray beam.**  
**Number (fraction) of potential defects in ROI:** 0 of 113569 (0.0E+00)  
**Number (fraction) of zero dark pixels in ROI:** 0 of 113569 (0.00)  
**Number of unconnected potential defects in ROI:** 0


**Potential-Defect Size Histogram**




**Pixel-Noise Histogram**




**Raw Dark Histogram**



**Raw Open Histogram**



**Good-Pixel Dark Histogram**



**Good-Pixel Open Histogram**

